

A-A-51282
9 December 1985

COMMERCIAL ITEM DESCRIPTION

PAD, HOSPITAL STRETCHER

The General Services Administration has authorized the use of this commercial item description in preference to MIL-P-37735.

This commercial item description covers a hospital stretcher pad intended for use with a wheeled hospital stretcher.

Salient characteristics:

The pad shall be a 2 inch thick foam pad with a removable cover that meets the conductivity requirements of NFPA No. 100. The cover material shall have a thickness of .015 to .020 inches (0.381 to 0.508 mm). The pad shall be approximately 25 inches (635 mm) wide and 76 inches (1930 mm) long. Each pad shall bear the manufacturer's name or registered trademark. The foam pad material shall be either vinyl, urethane, or any suitable synthetic rubber. The cover material shall be a suitable natural or synthetic rubber or vinyl treated for conductivity.

Workmanship. The hospital stretcher pad shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One pad, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release, distribution is unlimited.

AMSC N/A

A-A-51282

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-063-3967

This pad is intended for use on wheeled stretcher NSN 6530-01-063-3966.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

PREPARING ACTIVITY:

Custodians:

DoD - MB

Army - MD

Agent:

Navy - MS

Air Force - 03

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA - OSS

Project No. 6530-1823

Location H1-17-17