

GG-T-321D
AMENDMENT-2
July 17, 1978
SUPERSEDING
Amendment-1
June 12, 1963

FEDERAL SPECIFICATION

THERMOMETERS, SELF-INDICATING,
LIQUID-IN-GLASS FOR MACHINERY AND PIPING SYSTEMS

This amendment was developed by the General Services Administration, Federal Supply Service, Washington, DC 20406, based upon currently available technical information. It is recommended the Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

The General Services Administration has authorized the use of this amendment as a valid exception to Federal Specification GG-T-321D, dated July 11, 1977.

PAGE 6

3.3.1.1 Delete the last sentence.

PAGE 8

4.2.3 Delete paragraph and substitute:

"4.2.3 Sampling for examination and interchangeability. Sample thermometers and wells shall be selected from each lot in accordance with MIL-STD-105. The applicable AQL shall be 6.5 percent defective, and the inspection level shall be S-3."

4.2.4 Delete paragraph and substitute:

"Sampling for lot acceptance tests. A random sample of thermometers and wells shall be selected from each lot in accordance with MIL-STD-105, and shall be subjected to the tests specified in 4.4.1 through 4.4.5 to determine conformance to the requirements of this specification. The applicable AQL shall be 1.5 percent defective and the inspection level shall be S-2."

4.3.1, lines 4 and 5. Delete the sentence "The applicable AQL shall be 6.5 percent defective."

FSC 6685