

MIL-STD-1424

1 NOVEMBER 1979

MILITARY STANDARD

HYDROGEN ION METERS



FSC 6630

DEPARTMENT OF DEFENSE
Washington, DC 20301

Hydrogen Ion Meters

MIL-STD-1424

1. This Military Standard is approved for use by all Departments and Agencies of the Department of Defense.
2. Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of any use in improving this document should be addressed to: Commander, US Army Armament Research and Development Command, Attn: DRDAR-TSC-S, Aberdeen Proving Ground, MD 21010, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FOREWORD

This is the first book format standard generated on hydrogen ion meters. This standard is mandatory for use by all departments and agencies of the Department of Defense in the selection of items for application. It is intended to prevent the entry of unnecessary items (sizes, types and varieties) into the Department of Defense logistics system. This is not a procurement document.

CONTENTS

Paragraph		Page
1	SCOPE-----	1
1.1	Coverage-----	1
1.2	Application-----	1
2	REFERENCED DOCUMENTS-----	1
3	GENERAL REQUIREMENTS-----	2
3.1	Chemical and physical requirements-----	2
3.2	Nomenclature-----	2
3.3	Packaging data and labeling-----	2
3.4	Safety-----	2
3.5	Temperature-----	2
3.6	Use data-----	2
4	DETAIL REQUIREMENTS-----	2
4.1	General purpose hydrogen ion meters-----	2
4.1.2	Technical description-----	2
	Table I-----	3
4.1.3	Use data-----	4
4.1.4	Packaging data and labeling-----	4
4.1.5	Safety-----	4
4.1.6	Storage-----	4
4.2	pH meters for research-----	4
4.2.1	Specifications-----	4
4.2.2	Technical description-----	4
	Table II-----	5
	Table III-----	7
4.2.3	Use data-----	8
4.2.4	Packaging data and labeling-----	8
4.2.5	Safety-----	8
4.2.6	Storage-----	8

1. SCOPE

1.1 Coverage. This standard is a presentation of nomenclature, symbols, physical and chemical properties and requirements, military and typical commercial uses, directions for use, packaging data, toxicity data and shelf life of all military standard hydrogen ion meters. This standard does not necessarily include all classifications of the items represented by the title or those which are commercially available. It does contain items preferred for use in the selection of hydrogen ion meters for application by the Department of Defense. This standard covers the following 54 items:

<u>NAME</u>	<u>NO. OF ITEMS</u>
GENERAL PURPOSE pH METERS	37
pH METERS FOR RESEARCH	17

1.2 Application. Items listed herein accomodate essential requirements of the military and defense agencies, and will effect continued economies in all logistics functions when properly employed in new applications.

2. REFERENCED DOCUMENTS

2.1 Issues of documents. The issue of the following documents in effect on the date of invitation for bids form a part of this standard to the extent specified herein.

Federal Specification

PPP-B-636 Box, Shipping, Fiberboard

Military Specification

MIL-M-36755 Meters, Hydrogen Ion Test

Military Standard

MIL-STD-129 Marking for Shipment and Storage

2.2 Other publications. The following documents form a part of this standard to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

ANSI Standards

C39.5 Electrical and Electronic Measuring and Controlling
Instrumentation, Safety Requirements for

(Applications for copies should be addressed to the American Standards Institute, Incorporated, 1430 Broadway, New York, NY 10018.)

Clinical Laboratory Standards

PSI-5 Power Requirements for Clinical Laboratory Instruments
and Laboratory Power Lines

ASI-1 Preparation of Manual for Installation, Operation
and Repair of Laboratory Instruments

(Applications for copies should be addressed to the National Committee for
Clinical Laboratory Standards, 771 East Lancaster Avenue, Villanova, PA 19085.)

Underwriters Laboratories

UL1262 Laboratory Equipment

(Applications for copies should be addressed to Underwriters Laboratories,
Incorporated, 207 East Ohio Street, Chicago, IL 60611.)

3. GENERAL REQUIREMENTS

3.1 Chemical and physical requirements. All values given in tables of
chemical and physical requirements are in percent by weight unless otherwise
indicated.

3.2 Nomenclature. The Department of Defense item names, as used throughout
this standard, are in capital letters. Other names that are sometimes used
commercially are in small letters immediately beneath.

3.3 Packaging data and labeling. All meters in this standard shall be
packaged in accordance with Federal Specification PPP-B-636 and all applicable
documents mentioned in this military standard.

3.4 Safety. All general purpose hydrogen ion meters should be designed to
comply with ANSI C39.5, UL1262, and clinical standards PSI-5 and ASI-1
standards for safety and laboratory instruments. In using these meters
normal laboratory safety precautions should be followed.

3.5 Temperature. If the temperature at which a property was determined is
not specified, it is to be room temperature (20 to 25°C or 68 to 77°F).

3.6 Use data. Typical commercial uses are given. In general hydrogen
ion meters are designed for use in pH determinations and potentiometric
titrations.

4. DETAIL REQUIREMENTS

4.1 Name. GENERAL PURPOSE HYDROGEN ION METERS
General Purpose pH Meters

4.1.1 Specifications. MIL-M-36755, Meters, Hydrogen Ion Test.

4.1.2 Technical description. Table I cites the physical requirements of
general purpose pH meters. General purpose hydrogen ion meters are designed to
meet the requirements of industrial and educational laboratories. In addition
to doing routine pH measurements, general purpose pH meters usually have a
millivolt scale so that they can be used for potentiometric titrations. General
purpose pH meters, having a millivolt scale can be used for Karl Fisher titrations,
specific ion determinations and other dead stop titrations. Many of these meters
can be operated both on line and/or by batteries, offering the portability
necessary for field operations.

TABLE I. General Purpose Hydrogen Ion Meters

Meter	1	2	3	4	5
Ranges	0-14 pH ^a ± 700 mV ± 1400 mV	0-14pH ^a ± 1999 mV	0-14 pH ----	0-14 pH ± 1800 mV	2-12 pH ----
Smallest graduation nominal	0.10 pH ^b 10.0 mV	0.01 pH 1.0 mV	0.10 pH ----	----	0.20 pH ----
Temperature compensator	included manual or automatic	included manual or automatic	optional	included manual or automatic	optional
Temperature range	0-100°C	0-100°C	0-100°C	0-100°C	0-100°C
Installation	optional	optional	optional	optional	portable
Repeatability nominal	0.02 pH ^c 2.0 mV	0.01 pH 1.0 mV	0.02 pH ----	0.01 pH 1.0 mV	0.1 pH ----
Accuracy nominal	0.05 pH ^d 5.0 mV	0.01 pH ----	0.05 pH ----	0.02 pH 1.0 mV	----
Power	120/240 V 50/60 Hz	120/220 V 50/60 Hz	220/240 V 95/125 V 50/60 Hz	115 V 50/60 Hz	rechargeable battery pack

^a Also available in digital models

^b Also available in graduations of 0.05 pH and 5 mV

^c Also available with repeatability of 0.01 pH and 0.6 mV

^d Also available with an accuracy of 0.1 pH and 15 mV and 1.0 mV

4.1.3 Use data. General purpose hydrogen ion meters are intended for use in the routine determination of pH of solutions. If the pH meter is equipped with a millivolt scale, it can be used for potentiometric titrations, Karl Fisher titrations and other routine determinations.

4.1.4 Packaging data and labeling. For military use general purpose hydrogen ion meters shall be packed in accordance with PPP-B-636. Each unit package shall be marked in accordance with MIL-STD-129.

4.1.5 Safety. All general purpose hydrogen ion meters should be designed to comply with UL1262 and ANSI C39.5 standards for safety in laboratory instruments.

4.1.6 Storage. General purpose hydrogen ion meters should be stored in a cool, dry area.

4.2 Name. HYDROGEN ION METERS FOR RESEARCH
Ph Meters for Research

4.2.1 Specifications. None

4.2.2 Technical description. Table II cites the physical requirements of pH meters for research. pH meters for research are designed to give the precision necessary for research work. In addition many pH meters for research can perform potentiometric and amperometric titrations and specific ion concentration determinations. Many of these meters can be operated both on line and/or by batteries, offering the portability necessary for field operations.

TABLE II. Hydrogen Ion Meters for Research

Meter	1	2
Ranges	Normal 0-14 pH ± 700 mV	Expanded any 2.8 pH units
Smallest graduation	0.1 pH 10 mV	0.02 pH
Repeatability	± 0.02 pH ± 2 mV	0.002 pH ---
Accuracy	± 0.02 pH ± 4 mV	0.004 pH ---
Temperature compensator	optional	included manual or automatic
Temperature range	0 - 100°C	0 - 100°C
Installation	portable	optional
Power	batteries or 90/250 VAC 50/60 Hz	120/240 V 50/60 Hz
Differentiating meter characteristics	1. Built in battery charger 2. rugged splash proof case	1. has a log scale 2. provides test button for proper meter operation

TABLE II. (Continued)

Meter	3 Expanded Value	4 Normal	Expanded
Ranges	0-14 pH ± 1400 mV	0-14 pH ± 1400 mV	any 2 pH ± 200 mV
Smallest graduation	0.002 pH 0.2 mV	0.1 pH 10 mV	0.01 pH 1 mV
Repeatability	± 0.002 pH ± 0.2 mV	± 0.02 pH ± 2 mV	± 0.002 pH ± 0.2 mV
Accuracy	± 0.002 pH when standardized within 3 pH units 1/4 full mV scale	± 0.05 pH within 3 pH units ± 7 mV	± 0.007 pH ± 0.7 mV
Temperature compensator	included manual or automatic	included manual or automatic	
Temperature range	0 - 100°C	0 - 100°C	
Installation	fixed	fixed	
Power	90/130 V 50/60 Hz	120/240 V 50/60 Hz	
Differentiating meter characteristics	<ol style="list-style-type: none"> 1. pH range expands without manual switching 2. all solid state circuitry 3. housing is compact, rugged metal case with chemical resistant epoxy finish 	<ol style="list-style-type: none"> 1. has two and four decade log scales 2. can be used to measure monovalent and divalent ions 	

TABLE III. Digital Hydrogen Ion Meters for Research

Model	1	2	3
Ranges	Continuously Expanded 0-14 pH ± 1400 mV	0-14 pH ± 1999.9 mV	0-13.999 pH ± 1999.9 mV
Repeatability	± 0.001 pH ± 0.1 mV	± 0.001 pH ± 0.1 mV	N.A.*
Accuracy	± 0.002 pH ± 0.2 mV	± 0.001 pH N.A.	± 0.002 pH ± 0.1 mV or 0.05% of reading, whichever is greater
Temperature compensator	included automatic or manual	included automatic or manual	included automatic or manual
Temperature range	0 - 100°C	0 - 100°C	0 - 100°C
Installation	optional	optional	optional
Power	120/240 V 50/60 Hz.	120/240 V 50/60 Hz	100/115/220/240 V ± 10% 50/60 Hz
Differentiating meter characteristics	1. automatically indicates changes in polarity 2. has self-test circuit which indicates instrument malfunction 3. has analog and binary coded digital output for computer or printer	1. has electronic test switch to check for proper operation 2. has a blank switch that eliminates 3rd decimal place when not needed 3. has a multiple electrode selectrode capability	1. has an out of range indicator 2. has a binary coded digital output for digital printer or computer

*Not available

MIL-STD-1424

4.2.3 Use data. Expanded range hydrogen ion meters are intended for use in pH determinations where very precise determinations are necessary. Also these meters are used in potentiometric and redox titrations and in specific ion concentration determinations.

4.2.4 Packaging data and labeling. For military use expanded range hydrogen ion meters shall be packed in accordance with PPP-B-636. Each unit package shall be marked in accordance with MIL-STD-129.

4.2.5 Safety. All general purpose hydrogen ion meters should be designed to comply with ANSI C39.5 and UL1262 standards for safety in laboratory instruments.

4.2.6 Storage. Expanded range hydrogen ion meters should be stored in a cool, dry area.

Custodians:

Army - MD

Navy - MS

Air Force - 03

Other - DLA-DM

Preparing Activity: Army - EA

Project Number: 6630-0314

☆ U.S. GOVERNMENT PRINTING OFFICE: 1979 — 603-121/245

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER (Number) AND TITLE

MIL-STD 1424

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

VENDOR USER MANUFACTURER

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY (Printed or typed name and address - Optional)

TELEPHONE NO.

DATE

DD FORM 1426
1 OCT 76

EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

FOLD

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE \$300

POSTAGE AND FEES PAID

DOD 314



Commander
US Army Armament Research and Development Command
Attn: DRDAR-TSC-S
Aberdeen Proving Ground, MD 21010

FOLD