

MIL-STD-1665  
NOTICE 6  
23 April 1985

MILITARY STANDARD

TEST EQUIPMENT FOR THE  
STANDARD ELECTRONIC MODULES PROGRAM

TO ALL HOLDERS OF MIL-STD-1665:

1. THE FOLLOWING PAGES OF MIL-STD-1665 ARE NEW OR HAVE BEEN REVISED AND SUPERSEDE THE PAGES LISTED:

NEW PAGE	DATE	SUPERSEDED PAGE	DATE
5	23 April 1985	5	1 May 1984
6	1 May 1984	6	Reprinted without change
805-1	23 April 1985		
805-2	23 April 1985		

2. RETAIN THIS NOTICE AND INSERT BEFORE TABLE OF CONTENTS.

3. Holders of MIL-STD-1665 will verify that additions indicated above have been entered. This notice will be retained as a check sheet. This issuance, together with the appended pages is a separate publication. Each notice is to be retained by stocking points until the Military Standard is completely revised or canceled.

Custodians:  
Army - ER  
Navy - EC  
Air Force - 85

Preparing activity:  
Navy - EC  
(Project 5963-0034)

Review activities:  
Army - AT, AV  
Navy - AS, SH  
Air Force - 13, 17, 19  
DLA - ES

User activities:  
Army -  
Navy -  
Air Force -

Agent:  
DLA - ES

## INDEX

<u>Equipment</u>	<u>Item No.</u>
Ammeter, dc - - - - -	X01
Ammeter, dc - - - - -	X02
Ammeter, dc - - - - -	X03
Ammeter, dc - - - - -	X04
Ammeter, dc - - - - -	X05
Amplifier, low noise- - - - -	A03
Amplifier, power- - - - -	A01
Analyzer, distortion- - - - -	B01
Analyzer, signal- - - - -	B05
Analyzer, spectrum- - - - -	B03
Analyzer, wave- - - - -	B02
Analyzer, wave- - - - -	B04
Attenuator- - - - -	R08
Attenuator, balanced- - - - -	R09
Bridge, capacitance - - - - -	C01
Bridge, capacitance - - - - -	C03
Bridge, capacitance - - - - -	C04
Bridge, capacitance, rf - - - - -	C02
Bridge, inductance- - - - -	L01
Bridge, inductance- - - - -	L02
Bridge, inductance, rf - - - - -	L03
Bridge, megohm- - - - -	R06
Bridge, resistance- - - - -	R02
Bridge, resistance, rf- - - - -	R03
Counter-timer - - - - -	T01
Decade, synchro/resolver standard - - - - -	F03
Digital multifunction meter - - - - -	M02
Filter, tunable - - - - -	N01
Frequency meter - - - - -	T02
Frequency meter - - - - -	T03
Generator, function - - - - -	G05
Generator, function - - - - -	G06
Generator, function - - - - -	G10
Generator, pulse- - - - -	J01
Generator, pulse- - - - -	J02
Generator, pulse- - - - -	J04
Generator, pulse- - - - -	J05
Generator, pulse- - - - -	J06
Generator, power- - - - -	Q01
Generator, random noise - - - - -	K03
Generator, rf signal- - - - -	G11
Generator, sine wave- - - - -	G01
Generator, sine wave- - - - -	G02
Generator, sine wave- - - - -	G03
Generator, sine wave- - - - -	G04
Generator, sine wave- - - - -	G08
Generator, power sine wave- - - - -	G07
Generator, power sine wave- - - - -	G09
Generator, square wave- - - - -	H01
Generator, square wave- - - - -	H02
Generator, time-mark- - - - -	K01
Generator, word - - - - -	J03
Leakage tester, high voltage dc - - - - -	U01
Leakage tester, high voltage ac - - - - -	U02
Load, electronic- - - - -	EL01
Meter, capacitance- - - - -	C05
Multifunction meter, analog - - - - -	M03
Multimeter (VOM)- - - - -	M01
Ohmmeter- - - - -	R01
Ohmmeter- - - - -	R04
Ohmmeter- - - - -	R05
Ohmmeter- - - - -	R07
Ohmmeter- - - - -	R10

Supersedes page 5 of  
1 May 1984

## NOTICE 6

## INDEX - Continued

<u>Equipment</u>	<u>Item No.</u>
Oscilloscope, high frequency - - - - -	D01
Oscilloscope, high frequency - - - - -	D03
Oscilloscope, high frequency - - - - -	D06
Oscilloscope, high frequency - - - - -	D07
Oscilloscope, high gain - - - - -	D02
Oscilloscope, UHF sampling system - - - - -	D04
Oscilloscope, UHF sampling system - - - - -	D05
Phasemeter - - - - -	Z03
Photometer, digital - - - - -	K02
Power supply, constant current - - - - -	P05
Power supply, dc - - - - -	P01
Power supply, dc - - - - -	P02
Power supply, dc - - - - -	P03
Power supply, dc - - - - -	P04
Power supply, dc - - - - -	P06
Power supply, dc - - - - -	P07
Probe, oscilloscope current - - - - -	Y03
Probe, oscilloscope current - - - - -	Y05
Probe, oscilloscope voltage - - - - -	Y02
Probe, oscilloscope voltage - - - - -	Y04
Probe, oscilloscope voltage - - - - -	Y06
Probe, wide band - - - - -	Y01
Resolver/synchro bridge - - - - -	F04
Standard, dc voltage - - - - -	P08
Standard, dc voltage and null meter - - - - -	P09
Tester, LSI - - - - -	U03
Transformer, isolation - - - - -	E01
Transformer, isolation - - - - -	E02
Transformer, isolation - - - - -	E03
Transformer, isolation - - - - -	E05
Transformer, ratio - - - - -	F01
Transformer, ratio - - - - -	F02
Transformer, ratio - - - - -	F05
Transformer, Scott "T" - - - - -	E04
Voltmeter, ac - - - - -	W01
Voltmeter, ac - - - - -	W02
Voltmeter, ac - - - - -	W03
Voltmeter, ac - - - - -	W04
Voltmeter, ac - - - - -	W05
Voltmeter, ac - - - - -	W06
Voltmeter, dc - - - - -	V01
Voltmeter, dc - - - - -	V02
Voltmeter, dc - - - - -	V03
Voltmeter, dc - - - - -	V04
Voltmeter, dc - - - - -	V05
Voltmeter, dc - - - - -	V06
Voltmeter, differential ac/dc - - - - -	V07
Voltmeter, phase angle - - - - -	Z01
Voltmeter, phase angle - - - - -	Z02
Voltmeter, rf - - - - -	W08
Voltmeter, true rms - - - - -	W07
Wattmeter, ac - - - - -	S01
Wattmeter ac/dc - - - - -	S02

ITEM 805  
ANALYZER, SIGNAL

TYPE OF EQUIPMENT - - - - - Automatic signal analyzer.

FUNCTION PERFORMED - - - - - Performs precise frequency and amplitude measurements in the 10 Hz to 13 MHz region.

## SPECIFICATIONS

Frequency characteristics	
Range	10 Hz to 13 MHz.
Scanwidth	Any desired scan in 10,100, or 1000 steps of frequency increments as small as 0.1 Hz with a 0.1 Hz resolution.
Resolution	
Bandwidth	3 Hz to 10 kHz in a 1, 3, 10 sequence.
Bandwidth selectivity	60 dB/3 dB bandwidth ratios $\leq$ 11:1.
Stability	
Long term	$\pm 1 \times 10^{-8}/\text{day}$ ; $\pm 1 \times 10^{-6}/\text{month}$ .
Temperature	
20°C to 30°C	$\pm 1 \times 10^{-8}$ of frequency at 25°C.
0°C to 55°C	$\pm 1 \times 10^{-7}$ of frequency at 25°C.
Phase noise	<50 dB down in a 30 kHz band.
Center frequency accuracy	
3 Hz to 100 Hz, 1 kHz, 3 kHz bandwidth	$\pm 10\%$ of bandwidth setting.
300 Hz bandwidth	$\pm 45$ Hz.
10 kHz bandwidth	$\pm 500$ Hz.
Amplitude specifications	
Absolute amplitude calibration ranges	-130 dBm to +20 dBm (50 or 75 Ω) -140 dBV to +10 dBV.
Digital amplitude readout	$\pm 199.99$ dB with 0.01 dB resolution.
Dynamic range	
Average noise level	-127 dBV in 1 kHz resolution bandwidth.
Spurious responses	>70 dB below input range setting.
Distortion responses	>80 dB below input signal at input range setting level.
Power line related responses	70 dB below input range on +10 dBV through -40 dBV ranges; 60 dB on -50 dBV; 50 dB on -60 dBV ranges.

MIL-STD-1665  
NOTICE 6

B05 - Continued

**Amplitude accuracy**

Frequency response - - - - - ±0.25 dB (250 kHz reference).

Input range - - - - - ±0.05 dB/step; ±0.15 dB total accumulation.

**Log linearity**

0 to -30 dB - - - - - ±0.1 dB.

-30 to -60 dB - - - - - ±0.25 dB.

-60 to -80 dB - - - - - ±0.75 dB.

Input power - - - - - 115 or 230V ±10%, 48 to 63 Hz.

**EQUIPMENT MEETING ALL SPECIFICATIONS**

Manufacturer - - - - - Hewlett Packard Model 3044A Signal Analyzer.