

REVIEW ACTIVITY: ARMY - MD

FED. SUP CLASS  
6640

PART NUMBER	DIAMETER mm	PACKAGE SIZE lb
MS36259-3	0.177-0.250	1
*MS36259-1	0.5	5
*MS36259-2	5	1

\* Nominal Diameter

MATERIAL: MS36259-1 shall be of barium crown glass.

MS36259-2 and MS36259-3 shall be of borosilicate glass in accordance with Federal Specification DD-G-541, Type I, Class A.

DESIGN: Beads shall be solid spheres.

- NOTES:
1. Procurement document for MS36259-1 is Military Specification MIL-B-36282.
  2. Procurement document for MS36259-2 is Military Specification MIL-B-3473.
  3. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.
  4. For design feature purposes this standard takes precedence over procurement documents referenced herein.
  5. Referenced documents shall be of the issue in effect on date of invitations for bid.

(A) ENTIRE STANDARD REVISED

P.A.	MU	TITLE	MILITARY STANDARD
Other Cust	MS	BEAD, GLASS, LABORATORY	MS 36259
Procurement Specification See Notes 1 & 2	SUPERSEDES:		SHEET 1 OF 1

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

DD FORM 672 (Coordinated)  
1 SEP 65

EDITION OF 1 SEP 67 MAY BE USED

APPROVED 15 Dec 1961  
REVISED (A) 12 MAR 1969