

ARMY - MD, DM  
NAVY - MS

USER ACTIVITIES:


ARMY - ME  
AIR FORCE - 82

REVIEWER ACTIVITIES:

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

FED. SUP CLASS

6630



PART NUMBER	CAPACITY mL	HEIGHT mm
MS51439-1	125	118
MS51439-2	250	150
MS51439-3	500	170
MS51439-4	1000	205

**MATERIAL:** Bulb shall be made of borosilicate glass in accordance with Federal Specification DD-G-541, type I.

**DESIGN:** Bulb shall have a thick wall for mechanical strength. All bulbs shall have the same size tubulation to accommodate 6.4 mm rubber tubing.

**NOTE:** Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

APPROVED 5 MAY 1983 REVISED

P.A.	EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust	MS 99		BULB, LEVELING, GAS ANALYSIS APPARATUS	MS 51439
Procurement Specifications	NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 672 (COORDINATED)  
1 MAY 73

PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6630-0351