

ARMY - MD, DM  
NAVY - MS

USER ACTIVITIES:

ARMY - ME  
AIR FORCE - 82

REVIEWER ACTIVITIES:

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

FED. SUP CLASS

6630



PART NUMBER	CAPACITY mL	SUBDIV mL
MS51440-1	25	0.10
MS51440-2	50	0.10
MS51440-3	100	0.20
MS51440-4	250	1.00

**MATERIAL:** Burets shall be made of borosilicate glass in accordance with Federal Specification DD-G-541, type I.

**DESIGN:** Burets shall be made of straight tubing, the delivery end fire polished and tooled to proper size for connection with a delivery tip by means of flexible tubing. The scale shall be permanently etched and colored black. Every tenth graduation on a single scale shall be numbered to deliver.

**NOTE:** Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

P.A.	EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust	MS 99		BURET, LABORATORY	MS51440
Procurement Specifications	NONE	SUPERSEDES:	SHEET	OF
			1	1

DD FORM 672 (COORDINATED)

PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE PROJECT NUMBER 6630-0351

APPROVED 5 MAY 1983 REVISED